

WRIST STRAP & SHOE TESTER

61-200-0100

Sophisticated tester, designed for checking the conformity of operators wrist strap and/or shoes to the ESD protection standard. It is also designed to be permanent wall mounted in a production facility for daily testing of employee's wrist straps and footwear.

Product specifications:

The kit includes:

Tester

- Dimension: 150x88x35mm
- Weight: 220 grams
- Display: 1 green led / 6 red led / 1 vibror
- Voltage measurement: Measurement under permanent voltage of 20V +/- 5%

Shoe electrode

- Dimension: 370x440x50mm
- Weight: 700 grams
- Feature: double shoe electrode

Power supply

- Batteries: Supplied with 4 AA batteries
- Option: AC adapter



Calibration:

The tester needs to be calibrated yearly.

Possibility to calibrate it in-house with calibration plug (100kΩ - 35MΩ)

Other limits available on request

Options:

- | | |
|----------------------------------|-------------------------------|
| Power supply (European plug): | 61-200-0103 |
| Power supply (UK plug): | 61-300-0103 |
| Wall mounted operational panel: | 61-200-0105 (English version) |
| Wall mounted operational panel: | 61-200-0104 (French version) |
| PC interface: | 61-200-0107 |
| Calibration plug (100kΩ - 35MΩ): | 61-200-0112 |